

Certificate of Compliance

836 North Street - Life Sciences Tewksbury MA 01876 USA www.corning.com/lifesciences Refer to website for regional contact information.

Page: 1 / 1

Product Name : 1595 BTL 2 1/2 GAL/9500ML

Catalog Number : 1595-2X Manufacture Date : 2021-11-10

Lot ID : 31421001

Expiration Date : Unconfirmed Validity

Quality Management System - Complies with the current version of the ISO 9001 Standard.

Glass Type - Borosilicate, Type I, Class A, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

Animal Content - Product does not contain materials of animal origin.

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass Packaging Inspection - Pass Dimensional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.

ANDREW J (ANDY) FRANKS CMO Quality Manager

Creation Date: 2022-02-25 21:46:44