

## Certificate of Analysis

<b>Description:</b>	<b>TRANSPORT MEDIUM</b>	<b>Catalog #:</b>	25-500-CI
	Ingredients: Hanks' balanced salt solution, 2% heat inactivated fetal bovine serum, 100 µg/ml gentamicin, 0.5 µg/ml amphotericin B	<b>Lot #:</b>	01023018
<b>Storage:</b>	<b>2-8°C</b> DO NOT FREEZE	<b>Manufacture Date:</b>	2023-01
		<b>Expiration Date:</b>	2024-01
		<b>Country of Manufacture:</b>	USA

**Notes:**

Mediatech, Inc. products are prepared by a validated aseptic sterile filtration process to ensure that all products comply with the industry's Sterility Assurance Level (SAL) of at least  $10^{-3}$  (i.e. demonstrating that a product produced in this manner has no more than 1 random contaminant per 1000 units).

This product was formulated using the final ingredient concentrations outlined in CDC's SOP#: DSR-052-01: Preparation of Viral Transport Media.

This product was tested in accordance with currently approved Mediatech, Inc. specifications and procedures. Testing procedures are maintained in compliance with the current versions of the USP and/or EP, where applicable.

For Research and Further Manufacturing Use. Not for direct use in humans or animals. Utilization of this product apart from the labeled intended use may be a violation of local and/or Federal Law.

TEST PARAMETERS	SPECIFICATION	RESULT
pH	7.1 – 7.5 @ 20-25°C	7.2 @ 22°C
Osmolality	275 – 305 mOsm/kg H2O	290 mOsm/kg H2O
USP Sterility	Pass	Pass

For further information, contact Corning Scientific Support at 800-492-1110 or [ScientificSupport@corning.com](mailto:ScientificSupport@corning.com)

Following signatures indicate the above material has met all quality specifications and has been reviewed by a Quality representative.

Written By/Date: Luci Videncova 30 Jan 2023

Reviewed By/Date: [Signature] 31 JAN 2023



Mediatech, Inc.  
A Corning Subsidiary  
9345 Discovery Boulevard  
Manassas, VA 20109  
1-800-492-1110  
[www.corning.com](http://www.corning.com)