

## CERTIFICATE OF ANALYSIS

**PRODUCT:** Corning® Matrigel® Basement Membrane Matrix, 10 ml vial  
**CATALOG NUMBER:** 354234

**LOT NUMBER:** 16323001

**SOURCE:** Engelbreth-Holm-Swarm (EHS) Mouse Tumor  
**FORMULATION:** Dulbecco's Modified Eagle's Medium with 50 µg/ml gentamycin  
Corning Matrigel Basement Membrane Matrix is compatible with all culture media  
**STORAGE:** Store at -20°C. Avoid multiple freeze-thaws. Do not store in frost-free freezer. **KEEP FROZEN.**

### QUALITY CONTROL:

Specification	Criteria	Result
Protein Concentration	Results obtained by Lowry method and represented in mg/ml.	9.9
Endotoxin	Endotoxin units (EU)/ml are measured by Limulus Amoebocyte Lysate assay.	<3.5
Gelling	Tested for ability to gel quickly and maintain this form with culture medium for a period of 14 days at 37°C.	PASS
Biological Activity	Biological activity is determined using a neurite outgrowth assay. Chick dorsal root ganglia are plated on a 1.0 mm layer of Corning Matrigel Matrix. Tested for a positive neurite outgrowth response after 48 hours without addition of nerve growth factor.	PASS
Sterility	Tested for the presence of bacteria, fungi and mycoplasma.	NEGATIVE
MAP Test	Mouse colonies screened for Sendai, MHV, PVM, TEMV/GDVII, Ectro, Polyoma, MRV/EDIM, LCM, MCMV, M.Ad, Reo, MPV, LDEV/LDHV, MTV, Hantaan, K, RCMV, CARB	NEGATIVE
PCR Test	Tumor source tested for <i>Mycoplasma spp.</i> , <i>Helicobacter</i> , LDEV/LDHV, Sendai, MHV, PVM, MMV/MVM, MPV, Reo (1, 2, 3), MRV/EDIM, Ectro, LCM, K, MTV, Polyoma, Hantaan, Seoul, M. Ad (1, 2), MCMV, Norovirus, TMEV/GDVII, KRV, Toolan's H-1, RCV/SDA Finished goods tested for LDEV/LDHV.	NEGATIVE

**SAFETY RECOMMENDATION:** Handle in accordance with good industrial hygiene and laboratory safety practices

Expiration Date: August 13, 2025

Quality Assurance Eugene B. Hartigan

Date July 17, 2023

Discovery Labware, Inc., Two Oak Park, Bedford, MA 01730, Tel: 1.978.442.2200 (U.S.)  
CLSTechServ@Corning.com www.corning.com/lifesciences

CORNING

**For Research Use Only. Not for use in diagnostic or therapeutic procedures.**

For a listing of trademarks, visit [www.corning.com/lifesciences/trademarks](http://www.corning.com/lifesciences/trademarks)

© 2013 Corning Incorporated

All other trademarks in this document are the property of their respective owners.