Corning Inco Life Scien Registere ISO 9001:2		Certificate of Compliance
Catalog Number:	3913	
Lot Number:	20711021	
Product Description:	Costar® 1 x 8 Stripwell ^{M} , non-treated, solid white plate, flat bottom, without lid	
Component Materials: Strips -	Virgin Polystyrene, meets <i>USP, Class VI</i> requirements for plastic containers and closures. White concentrate Polystyrene, Heavy metal free (meets <i>CONEG</i> req.) color concentrate.	
Holder -		

Binding Characteristics:

Each lot has been tested for the attribute of well to well protein binding utilizing a luminescent immunoassay to measure precision. Coefficient of variation must be $\leq 8\%$.

Optical Characteristics:

The product is made of opaque white polystyrene to minimize well to well crosstalk and background fluorescence and /or luminescence.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

Test/Procedure:	Results:
Visual Attributes	Pass
Packaging	Pass
Functional	Pass

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Expiration Statement:

This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.

Judith A. Libbs

Judith A. Gibbs **Quality Assurance Manager**

Rev No: 5