### Corning Incorporated Life Sciences

Registered ISO 9001:2008

# Certificate of Compliance

Catalog Number: 3913

Lot Number: 31206008

Product Description: Costar® 1 x 8 Stripwell™, non-treated, solid white plate, flat bottom, without lid

**Component Materials:** 

Strips - Virgin Polystyrene, meets USP, Class VI requirements for plastic containers and

closures. White concentrate

Holder - Polystyrene, Heavy metal free (meets *CONEG* reg.) color concentrate.

## **Binding Characteristics:**

Each lot has been tested for the attribute of well to well protein binding utilizing a luminescent immunoassay to measure precision. Coefficient of variation must be  $\leq 8\%$ .

#### **Optical Characteristics:**

The product is made of opaque white polystyrene to minimize well to well crosstalk and background fluorescence and /or luminescence.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

Test/Procedure:Results:Visual AttributesPassPackagingPassFunctionalPass

#### Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

#### **Expiration Statement:**

This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.

Judith A. Gibbs

**Quality Assurance Manager** 

xuduth A. Glibbb

Rev No: 5