

**Corning Incorporated
Life Sciences**

Registered
ISO 9001:2008

Certificate of Compliance

Catalog Number: 3914

Lot Number: 29713030

Product Description: Costar® 1x8 Stripwell™, non-treated, solid black plate, flat bottom, without lid

Component Materials:

- Strips - Medium Impact Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures. Black concentrate
- Holder - Polystyrene, Heavy metal free (meets *CONEG* req.) color concentrate.

Binding Characteristics:

Each lot has been tested for the attribute of well to well protein binding utilizing a fluorescent immunoassay to measure precision. Coefficient of variation must be $\leq 15\%$.

Optical Characteristics:

The product is made of opaque black polystyrene to minimize well to well crosstalk and background fluorescence and /or luminescence.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:


Test/Procedure:	Results:
Visual Attributes	Pass
Packaging	Pass
Functional	Pass

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Expiration Statement:

This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.



Judith A. Gibbs
Quality Assurance Manager

Rev No: 6