Corning Inc Life Sci Regist ISO 900	ences ered	Certificate of Compliance	
Catalog Number:	430512		
Lot Number:	34211002		
Product Description:	Corning® 5 cellulose ac	Corning® 500 mL bottle top filter, 33 mm neck size, cellulose acetate membrane, 0.45 μ m pore size	
Component Materials: Funnel & Funnel Lid Funnel Adapter & Hose Connector Membrane Gasket Adapter Plug	 closures. Virgin High D containers ar Cellulose Acc Polyethylene 	Virgin Polystyrene, meets <i>USP, Class VI</i> requirements for plastic containers and closures. Virgin High Density Polyethylene, meets <i>USP, Class VI</i> requirements for plastic containers and closures. Heavy metal free (meets <i>CONEG</i> req.) color concentrate. Cellulose Acetate, 0.45 µm pore size Polyethylene Foam liner Cellulose Acetate	

Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.* Sterility Assurance Level: SAL 10⁻⁵

Pyrogens:

The product has been tested and has met the criteria established in the *Guideline on Validation of the Limulus* Amebocyte Lysate Test as an End-Product Endotoxin Test for Human and Animal Parenteral Drugs, Biological Products and Medical Devices (FDA, December, 1987). Results: less than 0.1 EU/mL

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product complies with the latest revision of EMEA/410/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human veterinary medicine products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMEA/410/01.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

Test/Procedure:	Results:
Membrane Integrity Test	Pass
Visual Attributes	Pass
Packaging	Pass

Expiration Statement:

This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Revision Date: 5-8-12

Wondy Lotsch

Wendy Loitsch Quality Manager

Rev No: 9