

**Corning Incorporated  
Life Sciences**

Registered  
ISO 9001:2008

**Certificate of Compliance**

**Catalog Number:** 431328  
**Lot Number:** 19213004  
**Product Description:** Corning® CellBIND® surface, 175 cm<sup>2</sup> cell culture flask with bar code, angled neck, vent cap

- Component Materials:**
- Flask - Virgin Polystyrene, meets *USP Class VI* requirements for plastic containers and closures.
  - Cap - Virgin High Density Polyethylene, meets *USP, Class VI* requirements for plastic containers and closures. Heavy metal free (meets *CONEG* req.) color concentrate. 0.2 µM microporous PTFE membrane, meets *USP, Class VI* requirements for plastic containers and closures.
  - Bar Code Label - White Polypropylene with adhesive, black bar code and characters.

**Bar Code Symbology:**  
Code 128 Subset A, 10 characters, human readable, starting with a "C"

**Sterilization:**  
This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.  
Sterility Assurance Level: SAL 10<sup>-6</sup>

**Pyrogens:**  
The product has been tested and has met the criteria established in the *Guideline on Validation of the Limulus Amebocyte Lysate Test as an End-Product Endotoxin Test for Human and Animal Parenteral Drugs, Biological Products and Medical Devices (FDA, December, 1987)*.  
Results: less than 0.1 EU/mL

**Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:**  
This product complies with the latest revision of EMEA/410/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human veterinary medicine products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMEA/410/01.

**Cell Attachment and Growth Characteristics:**  
The product has been tested for the attribute of Cell Attachment and Growth utilizing an attachment- dependent mammalian cell line in a serum supplemented media.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

<b>Test/Procedure:</b>	<b>Results:</b>
Leak Test	Pass
Visual Attributes	Pass
Cell Culture Treatment	Pass
Packaging	Pass

**Expiration Statement:**  
This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.

**Lot Number Designation:**  
8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

**Revision Date:**  
7-22-11

*Wendy Loitsch*

Wendy Loitsch  
Quality Manager

**Corning Incorporated  
Life Sciences**

Registered  
ISO 9001:2008

# Certificate of Compliance

**Catalog Number:** 431328

**Lot Number:** 19213004

**Product Description:** Corning® CellBIND® surface, 175 cm<sup>2</sup> cell culture flask with bar code, angled neck, vent cap

Rev No: 8