

836 North Street - Life Sciences  
Tewksbury MA 01876 USA  
www.corning.com/lifesciences  
Refer to website for regional contact information.

Page: 1 / 1

<b>Product Name</b>	: SQUARE GLASS FIBER PREFILTER, 48.5MM, FOR THE 250ML FUNNEL,		
<b>Catalog Number</b>	: 431411	<b>Manufacture Date</b>	: 2021-06-15
<b>Lot ID</b>	: 15921001		
<b>Expiration Date</b>	: Unconfirmed Validity		

**Quality Management System** - Complies with the current version of the ISO 9001 Standard.

**Quality Control Testing** - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

- Visual Inspection - Pass
- Packaging Inspection - Pass
- Dimensional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



ANDREW J (ANDY) FRANKS  
CMO Quality Manager