

Certificate of compliance

Date of certificate: 2020 - 07 - 02

Page: 1/1

CORNING GOSSELIN SAS 123 Rue de Caëstre - CS40019 59529 HAZEBROUCK CEDEX BORRE - FRANCE

Product description

Corning Octagonal PET Bottle 1000mL Graduated, 31mm IATA-Validated Screw Cap Sterile, Pre-assembled, 12/Tray, 24/Cas

Catalogue number 432334

Manufacturing date 2020/03/11

Lot number 741790

Raw material statement

The Substances Bisphenol A, BSE / TSE, latex, and phthalates are not used or intentionally incorporated in the manufacture of the raw materials or final product.

Sterility Assurance Level

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10-6

Leakproof - With a torque value of 2,5N.m (22lb.in), this product meets IATA Packaging Instruction 650 requirements to ensure leakproof performance during air transport conditions; however, this is not designed for packaging and transport of UN3373 Biological and Infectious Substances.

Non-Pyrogenic - Tested and met the criteria established in the current version of United States Pharmacopeia (USP) Chapter <85> "Bacterial Endotoxins Test." The acceptance level for product is less than 0.1 EU/ml.

USP Class VI Testing - All material resin is tested, qualified and shown to be non-toxic as established in the standards USP Class VI Chapter <87>, "Biological reactivity Tests, in Vitro" and Chapter <88>, "Biological Reactivity Tests, in Vivo."

Expiration date 2023/03/01

Expiration statement

This product is designed to have and expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level. We confirm that product is compliant with ISO993/USP31: Non-cytotoxic, Endotoxin-free / non-pyrogenic.

Cédric MAES

Plant Quality Manager, Borre

maesc@corning.com