

Corning Incorporated - Life Sciences
6507 South 400 West
Salt Lake City UT 84107 USA
www.corning.com/lifesciences
Refer to website for regional contact information.

Page: 1 / 1

Product Name	: CLUSTER TUBE,1.2ML,8 STRIP,PP,S,12BG/120CS		
Catalog Number	: 4418	Manufacture Date	: 2023-08-25
Lot ID	: 23723000		
Expiration Date	: 2026-08-25		

Quality Management System - Complies with the current version of the EN ISO 13485 Standard.

Sterility - Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products- Radiation". Products meet a minimum Sterility Assurance Level (SAL) of 10^{-3} .

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass
Packaging Inspection - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



Anthony Sloan
Quality Manager