

Corning Incorporated - Life Sciences
2 Alfred Road
Kennebunk ME 04043 USA
www.corning.com/lifesciences
Refer to website for regional contact information.

Page: 1 / 1

Product Name	: PIPETTE, 100ML, PPW, PS, S, IND, 10		
Catalog Number	: 4491	Manufacture Date	: 2022-07-10
Lot ID	: 19122011		
Expiration Date	: 2025-07-09		

Quality Management System - Complies with the current version of the ISO 9001 Standard.

Non-Pyrogenic - Tested and met the criteria established in the current version of ANSI/AAMI ST 72, "Bacterial Endotoxins - Test methodologies, routine monitoring, and alternatives to batch testing" and USP <85>, "Bacterial Endotoxins Test". The acceptance level for product is ≤ 0.10 EU/ml or ≤ 4 EU/device.

USP Class VI Testing - All material resin is tested, qualified and shown to be non-toxic as established in the Standards USP Class VI Chapter<87>, "Biological reactivity Tests, in Vitro" and Chapter<88>, "Biological Reactivity Tests, in vivo".

DNase/RNase Free - Tested by nuclease assay method and found to be free of detectable DNase/RNase contamination. The assay detection limit is 10^{-7} Kunitz units/uL for DNase and 10^{-9} Kunitz units/uL for RNase.

Human DNA Free - Tested by PCR method and found to be free of detectable human DNA contamination.

BSE-TSE - This product is deemed animal free by virtue of not containing materials of animal origin and/or complies with the latest revision of EMA/410/01 section 6.4.

Sterilization - Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products - Radiation". Products meet a minimum Sterility Assurance Level (SAL) of 10^{-6} .

Volumetric Accuracy - Serological pipets are accurate to +/- 2% at full volume in compliance with ASTM E934, "Standard Specification for Serological Pipet, Disposable Plastic" and ISO 12771, "Plastics laboratory ware - Disposable serological pipettes".

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

- Visual Inspection - Pass
- Packaging Inspection - Pass
- Integrity Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.

Olga I Stridiron Monell
Plant Quality Manager