

Corning Incorporated - Life Sciences
271 County Route 64
Big Flats NY 14814 USA
www.corning.com/lifesciences
Refer to website for regional contact information.

Page: 1 / 1

| | |
|---|--------------------------------------|
| Product Name : 1D/2D VIAL, 2.0ML, W/CAP, EXT, BK | |
| Catalog Number : 8671 | Manufacture Date : 2018-12-13 |
| Lot ID : 34718085 | |
| Expiration Date : 2024-01-04 | |

Quality Management System - Complies with the current version of the ISO 9001 Standard.

Animal Content - Product does not contain materials of animal origin.

Product Attribute - Validated for compliance to IATA PI 650; does not leak at a pressure differential of 95kPa in the range of -40°C to +55°C.

USP Class VI Testing - All material resin is tested, qualified and shown to be non-toxic as established in the Standards USP Class VI Chapter<87>, "Biological reactivity Tests, in Vitro" and Chapter<88>, "Biological Reactivity Tests, in vivo".

Non-Pyrogenic - Validated per European Pharmacopoeia criteria for pyrogens and meets the requirements for classification as non-pyrogenic. The acceptance level for product is ≤ 0.1 EU/mL.

DNase/RNase Free - Tested by nuclease assay method and found to be free of detectable DNase/RNase contamination. The assay detection limit is 10^{-6} Kunitz units/ul for DNase and 10^{-9} Kunitz units/ul for RNase.

Human DNA Free - Tested by PCR method and found to be free of detectable human DNA contamination. The assay detection limit is hDNA 2pg.

Sterilization - Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products - Radiation". Products meet a minimum Sterility Assurance Level (SAL) of 10^{-6} .

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release.

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



Edwin A Kindler
Div. Quality Engineering Manager