

CORNING GOSSELIN SAS
123 Rue de Caëstre - CS40019
59529 HAZEBROUCK CEDEX
BORRE - FRANCE
www.corning.com/lifesciences
Refer to website for regional contact information

Date of certificate : 2023 - 07 - 19

Page 1

Product description

Square Dish 120x15 4Vents Sterile 1 4/Bag 252/CS

Catalogue number

BP124-05

Lot number

859745

Manufacturing date

2023/06/01

Quality Management System

Complies with the current version of the ISO9001 standard.

Sterility Assurance Level

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10⁻⁶

Expiration date

2026/06/01

Expiration statement

This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release



Hugues Vanderschelden
Plant Quality Manager, Borre
vandersch2@corning.com