CORNING

CORNING GOSSELIN SAS 123 Rue de Caëstre - CS40019 59529 HAZEBROUCK CEDEX BORRE - FRANCE

Certificate of compliance

Date of certificate : 2018 - 11 - 14

Page : 1/1

Product description

Bottle RoundHDPE 100mL 37mmRedCapWad Assembled Sterile 335CS

Catalogue number LR100-12

Lot number 665152

Sanitary Assurance Level

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10-3

Manufacturing date 2017/10/25

Expiration date 2020/10

Expiration statement

This product is designed to have and expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

flan

Cédric MAES Plant Quality Manager, Borre & Chorges maesc@corning.com

Corning Restricted