

CORNING GOSSELIN SAS
123 Rue de Caëstre - CS40019
59529 HAZEBROUCK CEDEX
BORRE - FRANCE
www.corning.com/lifesciences
Refer to website for regional contact information

Date of certificate : 2023 - 07 - 18

Page 1

Product description

Bottle Oct PET 1L 31mmTECap Non-ass embled 12/Pack 48/CS

Catalogue number

P1000B-06

Manufacturing date

2022/07/15

Lot number

826505

Quality Management System

Complies with the current version of the ISO9001 standard.

Sterility Assurance Level

Non sterile

Expiration statement

N/A

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release



Hugues Vanderschelden
Plant Quality Manager, Borre
vandersch2@corning.com