

# Certificate of compliance

Date of certificate: 2022 - 10 - 21

Page 1

CORNING GOSSELIN SAS
123 Rue de Caëstre - CS40019
59529 HAZEBROUCK CEDEX
BORRE - FRANCE
www.corning.com/lifesciences
Refer to website for regional contact information

**Product description** 

Container 60mL PP RedScrewCap Steri le 70/Bag 700/CS

Catalogue number

TP35C-002

Lot number

827014

**Manufacturing date** 2022/07/01

Quality Management System: Complies with the current version of the ISO9001 standard

## **Sterility Assurance Level**

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10-3

#### **Expiration date**

2025/07/01

# **Expiration statement**

This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.

### **Performance testing (AQL)**

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release

Hugues Vanderschelden Plant Quality Manager, Borre vandersch2@corning.com