

CORNING GOSSELIN SAS  
123 Rue de Caëstre - CS40019  
59529 HAZEBROUCK CEDEX  
BORRE - FRANCE  
[www.corning.com/lifesciences](http://www.corning.com/lifesciences)  
Refer to website for regional contact information

Date of certificate : 2023 - 06 - 08

Page 1

### Product description

Container 60mL PP RedScrewCap Sterile 70/Bag 700/CS

### Catalogue number

TP35C-002

### Manufacturing date

2023/05/01

### Lot number

854497

### Quality Management System

Complies with the current version of the ISO9001 standard.

### Sterility Assurance Level

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10<sup>-3</sup>

### Expiration date

2026/05/01

### Expiration statement

*This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.*

### Performance testing (AQL)

*Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.*

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release



Hugues Vanderschelden  
Plant Quality Manager, Borre  
[vandersch2@corning.com](mailto:vandersch2@corning.com)