

CORNING GOSSELIN SAS
123 Route de Caëstre
59190 - Borre
FRANCE
www.corning.com/lifesciences
Refer to website for regional contact information

Date of certificate : 2023 - 09 - 13

Page 1

Product description

Container 180mL PP RedScrewCap Ster ile 264/CS

Catalogue number

TP53C-002

Manufacturing date

2023/07/01

Lot number

862326

Quality Management System

Complies with the current version of the ISO9001 standard.

Sterility Assurance Level

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10⁻³

Expiration date

2026/07/01

Expiration statement

This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release



Sandra Bianchi
Plant Quality Manager, Borre
BianchiS@corning.com